

CALIBRATION CERTIFICATE

LOGGER SERIAL NUMBER: 20687888

CALIBRATION RESULT: PASSED

Report Number	08_19_19_71
Certification Date	8/19/2019
Logger Type	HOBO Water Level Logger
Water Level Range	0 to 9 m (0 to 30 ft)
Logger Part Number	U20-001-01
Logger Status	New
Full Scale Pressure Range	0 to 207 kPa (0 to 30 psi)
Calibrated Range	69 to 207 kPa (10 to 30 psi), 0 to 40° C

Onset Computer Corporation certifies that the pressure accuracy of the data logger listed above has been observed to be within its published pressure specifications. Onset Computer's calibrated reference instruments are traceable to NIST, and certification files are maintained at Onset Computer's corporate headquarters in Bourne, MA.

Test Equipment and Procedures

Pressure Regulator and Calibrator: TE2-17200 (Calibrated on 07/23/2019)
 Environmental Chamber: TE1-17197
 Onset Calibration Software: D-17253
 Onset Calibration Procedure: D-17254
 Range of Applied Pressures: 69 to 207 kPa (10 to 30 psi)
 Range of Applied Temperature: 0-40 °C (Nominal)

Test Data

Pressure (psia)	
Applied	Observed
15.000	14.993
21.000	21.013
29.002	28.996

Results

Specified Absolute Pressure Accuracy: PASSED

Typical: ± 0.1 %FS, ± 0.21 kPa (± 0.03 psi)

Maximum: ± 0.3 %FS, ± 0.62 kPa (± 0.09 psi)

Test Performed By: Isaiah Andrea

This calibration report may not be reproduced, except in full, without the written approval of Onset Computer Corporation.

D-17507-B

Onset Computer Corporation 470 MacArthur Blvd. Bourne, MA 02532 USA

Tel: 508-759-9500 Fax: 508-759-9100 Web: www.onsetcomp.com Email: sales@onsetcomp.com